

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061424\
 Data File : VN082574.D
 Acq On : 14 Jun 2024 11:36
 Operator : JC\MD
 Sample : VN0614MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0614MBL01

Quant Time: Jun 14 23:43:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	285863	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	274895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114031	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	136322	54.700	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.400%
35) Dibromofluoromethane	8.171	113	98126	55.801	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.600%
50) Toluene-d8	10.571	98	367424	54.498	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.000%
62) 4-Bromofluorobenzene	12.853	95	121781	50.608	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	101.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061424\
Data File : VN082574.D
Acq On : 14 Jun 2024 11:36
Operator : JC\MD
Sample : VN0614MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0614MBL01

Quant Time: Jun 14 23:43:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

