

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061424\
 Data File : VN082582.D
 Acq On : 14 Jun 2024 14:50
 Operator : JC\MD
 Sample : P2891-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-380-382

Quant Time: Jun 14 23:46:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	156961	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	289649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	280070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	129908	52.415	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.820%	
35) Dibromofluoromethane	8.171	113	95207	53.433	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.860%	
50) Toluene-d8	10.571	98	350179	51.261	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.520%	
62) 4-Bromofluorobenzene	12.853	95	113868	46.701	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 93.400%	
Target Compounds						
16) Acetone	4.436	43	4092	4.025	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061424\
Data File : VN082582.D
Acq On : 14 Jun 2024 14:50
Operator : JC\MD
Sample : P2891-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-380-382

Quant Time: Jun 14 23:46:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

