

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061521\
 Data File : VN067353.D
 Acq On : 15 Jun 2021 18:01
 Operator : JC/MD
 Sample : M2678-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-AMND-COMP-K(CL-16)

Quant Time: Jun 16 02:40:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	351759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	666719	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	604149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	234192	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	330008	47.557	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.120%
35) Dibromofluoromethane	8.024	113	133478	30.102	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	60.200%#
50) Toluene-d8	10.443	98	856519	50.397	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.800%
62) 4-Bromofluorobenzene	12.734	95	330264	47.087	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.180%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	28850	10.834	ug/l	98
20) Methylene Chloride	5.095	84	30780	5.555	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061521\
Data File : VN067353.D
Acq On : 15 Jun 2021 18:01
Operator : JC/MD
Sample : M2678-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPLP-AMND-COMP-K(CL-16)

Quant Time: Jun 16 02:40:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

