

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061521\
 Data File : VN067354.D
 Acq On : 15 Jun 2021 18:26
 Operator : JC/MD
 Sample : M2678-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-AMND-COMPOSITE-H

Manual Integrations
 APPROVED

MMDadoda
 6/16/2021 5:03:25 PM

Quant Time: Jun 16 02:40:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	338939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	644197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	584729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	228148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	321766	48.123	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.240%		
35) Dibromofluoromethane	8.024	113	125877	29.380	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	58.760%#		
50) Toluene-d8	10.446	98	828605	50.459	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.920%		
62) 4-Bromofluorobenzene	12.736	95	323839	47.786	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	95.580%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	22504	8.770	ug/l	96
20) Methylene Chloride	5.095	84	22367	4.190	ug/l	97
25) 2-Butanone	7.359	43	6015m	1.641	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061521\
Data File : VN067354.D
Acq On : 15 Jun 2021 18:26
Operator : JC/MD
Sample : M2678-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPLP-AMND-COMPOSITE-H

Manual Integrations
APPROVED

MMDadoda
6/16/2021 5:03:25 PM

Quant Time: Jun 16 02:40:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

