

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061521\
 Data File : VN067357.D
 Acq On : 15 Jun 2021 19:38
 Operator : JC/MD
 Sample : M2709-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-AMND-COMP-M

Manual Integrations
 APPROVED

MMDadoda
 6/16/2021 5:03:26 PM

Quant Time: Jun 16 02:41:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	339578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	641236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	584151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	226461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	320966	47.913	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.820%		
35) Dibromofluoromethane	8.024	113	160329	37.594	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	75.180%		
50) Toluene-d8	10.446	98	831888	50.893	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.780%		
62) 4-Bromofluorobenzene	12.736	95	314735	46.657	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.320%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	23048	8.966	ug/l	93
20) Methylene Chloride	5.103	84	14730m	2.754	ug/l	
25) 2-Butanone	7.356	43	6759	1.840	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061521\
Data File : VN067357.D
Acq On : 15 Jun 2021 19:38
Operator : JC/MD
Sample : M2709-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPLP-AMND-COMP-M

Manual Integrations
APPROVED

MMDadoda
6/16/2021 5:03:26 PM

Quant Time: Jun 16 02:41:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

