

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078221.D
 Acq On : 15 Jun 2023 12:08
 Operator : JC\MD
 Sample : VN0615MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0615MBL01

Quant Time: Jun 16 06:03:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	426136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	774983	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	693983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	216856	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	305666	53.798	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.600%
35) Dibromofluoromethane	8.171	113	272577	55.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.260%
50) Toluene-d8	10.571	98	930968	49.282	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.560%
62) 4-Bromofluorobenzene	12.853	95	284282	48.408	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078221.D
Acq On : 15 Jun 2023 12:08
Operator : JC\MD
Sample : VN0615MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0615MBL01

Quant Time: Jun 16 06:03:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

