

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078225.D
 Acq On : 15 Jun 2023 13:56
 Operator : JC\MD
 Sample : 03231-04DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-1D-06132023DL

Quant Time: Jun 16 06:06:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	434685	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	795391	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	689950	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	212334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	305685	52.744	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.480%
35) Dibromofluoromethane	8.177	113	269604	53.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.220%
50) Toluene-d8	10.570	98	949012	48.949	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.900%
62) 4-Bromofluorobenzene	12.853	95	287570	47.712	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.420%
Target Compounds						Qvalue
44) Trichloroethene	9.359	130	272969	47.995	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078225.D
Acq On : 15 Jun 2023 13:56
Operator : JC\MD
Sample : 03231-04DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-1D-06132023DL

Quant Time: Jun 16 06:06:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

