

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061523\  
 Data File : VN078228.D  
 Acq On : 15 Jun 2023 15:08  
 Operator : JC\MD  
 Sample : 03259-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 GWOW-BR-05-167-187-061423

Quant Time: Jun 16 06:07:37 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060623W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 07 05:05:00 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.230	168	338300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	619334	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	543457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	151854	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	237992	52.763	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.520%	
35) Dibromofluoromethane	8.177	113	212334	54.223	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	108.440%	
50) Toluene-d8	10.571	98	778913	51.596	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.200%	
62) 4-Bromofluorobenzene	12.853	95	212467	45.272	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	90.540%	
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	4.347	96	470674	127.408	ug/l	90
16) Acetone	4.453	43	5011	3.860	ug/l #	74
21) trans-1,2-Dichloroethene	5.800	96	42485	10.663	ug/l	86
24) 1,1-Dichloroethane	6.577	63	58588	8.373	ug/l	97
27) cis-1,2-Dichloroethene	7.494	96	310886	65.412	ug/l	95
40) Benzene	8.612	78	5736	0.342	ug/l	100
42) 1,2-Dichloroethane	8.682	62	33899	5.896	ug/l	94
44) Trichloroethene	9.359	130	305527	68.991	ug/l	100
49) 1,4-Dioxane	9.694	88	7347	99.259	ug/l #	93
52) Toluene	10.635	92	53350	5.039	ug/l	99
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061523\  
 Data File : VN078228.D  
 Acq On : 15 Jun 2023 15:08  
 Operator : JC\MD  
 Sample : 03259-03  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 GWOW-BR-05-167-187-061423

Quant Time: Jun 16 06:07:37 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060623W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 07 05:05:00 2023  
 Response via : Initial Calibration

