

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078229.D
 Acq On : 15 Jun 2023 15:32
 Operator : JC\MD
 Sample : 03259-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-061423

Quant Time: Jun 16 06:08:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	311396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	579394	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	513925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	226843	54.636	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.280%	
35) Dibromofluoromethane	8.177	113	192303	52.493	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.980%	
50) Toluene-d8	10.576	98	709301	50.223	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	12.859	95	202782	46.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.380%	
Target Compounds						
					Qvalue	
44) Trichloroethene	9.359	130	2034	0.491	ug/l	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078229.D
Acq On : 15 Jun 2023 15:32
Operator : JC\MD
Sample : 03259-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-061423

Quant Time: Jun 16 06:08:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

