

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078233.D
 Acq On : 15 Jun 2023 17:09
 Operator : JC\MD
 Sample : 03231-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-8-06132023

Quant Time: Jun 16 06:09:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	417162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	771218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	677063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	210361	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	304128	54.679	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.360%
35) Dibromofluoromethane	8.177	113	260424	53.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.820%
50) Toluene-d8	10.576	98	907228	48.260	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.520%
62) 4-Bromofluorobenzene	12.853	95	277540	47.491	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.980%
Target Compounds						
30) Chloroform	7.977	83	4021	0.433	ug/l	86
38) Carbon Tetrachloride	8.359	117	4040	0.547	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078233.D
Acq On : 15 Jun 2023 17:09
Operator : JC\MD
Sample : 03231-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-8-06132023

Quant Time: Jun 16 06:09:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

