

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078237.D
 Acq On : 15 Jun 2023 18:45
 Operator : JC\MD
 Sample : 03231-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-EQB-2-06132023

Quant Time: Jun 16 06:11:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	420986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	776197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	695553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	222469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	308624	54.983	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.960%			
35) Dibromofluoromethane	8.177	113	274891	56.011	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 112.020%			
50) Toluene-d8	10.571	98	916929	48.463	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.920%			
62) 4-Bromofluorobenzene	12.853	95	290580	49.403	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.800%			
Target Compounds					Qvalue	
16) Acetone	4.448	43	10760	6.660	ug/l	93
29) Tetrahydrofuran	7.853	42	22004	13.248	ug/l #	85
30) Chloroform	7.977	83	235071	25.069	ug/l	96
40) Benzene	8.612	78	12290	0.585	ug/l #	80
47) Bromodichloromethane	9.894	83	9125	1.275	ug/l #	98
52) Toluene	10.635	92	3797	0.286	ug/l	98
67) Ethyl Benzene	11.971	91	7398	0.288	ug/l	100
68) m/p-Xylenes	12.071	106	8450	0.863	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078237.D
Acq On : 15 Jun 2023 18:45
Operator : JC\MD
Sample : 03231-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-EQB-2-06132023

Quant Time: Jun 16 06:11:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

