

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078239.D
 Acq On : 15 Jun 2023 19:34
 Operator : JC\MD
 Sample : 03231-19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-4D-06122023

Quant Time: Jun 16 06:11:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	417105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	775733	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	678304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	203978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	300153	53.972	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.940%
35) Dibromofluoromethane	8.177	113	270360	55.121	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.240%
50) Toluene-d8	10.571	98	916189	48.453	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.853	95	274308	46.665	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.320%
Target Compounds						
						Qvalue
44) Trichloroethene	9.359	130	149359	26.927	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078239.D
Acq On : 15 Jun 2023 19:34
Operator : JC\MD
Sample : 03231-19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-4D-06122023

Quant Time: Jun 16 06:11:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

