

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078240.D
 Acq On : 15 Jun 2023 19:58
 Operator : JC\MD
 Sample : 03231-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-4S-06122023

Quant Time: Jun 16 06:12:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	405059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	759317	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	675792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	201164	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	296935	54.981	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.960%	
35) Dibromofluoromethane	8.177	113	260274	54.212	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.420%	
50) Toluene-d8	10.571	98	904304	48.859	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.720%	
62) 4-Bromofluorobenzene	12.853	95	269966	46.919	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	2669	1.717	ug/l	95
44) Trichloroethene	9.359	130	390377	71.900	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078240.D
Acq On : 15 Jun 2023 19:58
Operator : JC\MD
Sample : 03231-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-4S-06122023

Quant Time: Jun 16 06:12:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

