

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078220.D
 Acq On : 15 Jun 2023 11:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 16 06:02:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.565	0.453	19.8	72	0.00
3 P	Chloromethane	0.611	0.506	17.2	76	0.00
4 C	Vinyl Chloride	0.740	0.645	12.8#	77	0.00
5 T	Bromomethane	0.559	0.484	13.4	76	0.00
6 T	Chloroethane	0.505	0.465	7.9	82	0.00
7 T	Trichlorofluoromethane	1.004	0.889	11.5	79	0.00
8 T	Diethyl Ether	0.353	0.318	9.9	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.530	0.483	8.9	80	0.00
10 T	Methyl Iodide	0.749	0.688	8.1	74	0.00
11 T	Tert butyl alcohol	0.092	0.077	16.3	73	0.00
12 CM	1,1-Dichloroethene	0.546	0.473	13.4#	75	0.00
13 T	Acrolein	0.101	0.081	19.8	70	0.00
14 T	Allyl chloride	0.667	0.696	-4.3	89	0.00
15 T	Acrylonitrile	0.237	0.233	1.7	80	0.00
16 T	Acetone	0.192	0.174	9.4	81	0.00
17 T	Carbon Disulfide	1.405	1.161	17.4	72	0.00
18 T	Methyl Acetate	0.883	0.834	5.5	83	0.00
19 T	Methyl tert-butyl Ether	1.742	1.669	4.2	80	0.00
20 T	Methylene Chloride	0.634	0.582	8.2	81	0.00
21 T	trans-1,2-Dichloroethene	0.589	0.531	9.8	77	0.00
22 T	Diisopropyl ether	1.568	1.602	-2.2	84	0.00
23 T	Vinyl Acetate	0.932	0.924	0.9	80	0.00
24 P	1,1-Dichloroethane	1.034	1.028	0.6	84	0.00
25 T	2-Butanone	0.295	0.284	3.7	79	0.00
26 T	2,2-Dichloropropane	0.764	0.833	-9.0	89	0.00
27 T	cis-1,2-Dichloroethene	0.702	0.644	8.3	79	0.00
28 T	Bromochloromethane	0.475	0.502	-5.7	88	0.00
29 T	Tetrahydrofuran	0.197	0.191	3.0	79	0.00
30 C	Chloroform	1.114	1.123	-0.8#	84	0.00
31 T	Cyclohexane	0.855	0.752	12.0	77	0.00
32 T	1,1,1-Trichloroethane	0.978	0.969	0.9	80	0.00
33 S	1,2-Dichloroethane-d4	0.667	0.694	-4.0	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.316	0.330	-4.4	84	0.00
36 T	1,1-Dichloropropene	0.453	0.446	1.5	80	0.00
37 T	Ethyl Acetate	0.354	0.350	1.1	80	0.00
38 T	Carbon Tetrachloride	0.479	0.477	0.4	81	0.00
39 T	Methylcyclohexane	0.470	0.387	17.7	68	0.00
40 TM	Benzene	1.354	1.345	0.7	81	0.00
41 T	Methacrylonitrile	0.192	0.191	0.5	85	0.00
42 TM	1,2-Dichloroethane	0.464	0.485	-4.5	87	0.00
43 T	Isopropyl Acetate	0.677	0.621	8.3	80	0.00
44 TM	Trichloroethene	0.358	0.341	4.7	79	0.00
45 C	1,2-Dichloropropane	0.337	0.349	-3.6#	85	0.00
46 T	Dibromomethane	0.248	0.245	1.2	82	0.00
47 T	Bromodichloromethane	0.461	0.504	-9.3	86	0.00
48 T	Methyl methacrylate	0.284	0.281	1.1	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
 Data File : VN078220.D
 Acq On : 15 Jun 2023 11:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 16 06:02:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	78	0.00
50 S	Toluene-d8	1.219	1.253	-2.8	83	0.00
51 T	4-Methyl-2-Pentanone	0.357	0.366	-2.5	82	0.00
52 CM	Toluene	0.855	0.853	0.2#	80	0.00
53 T	t-1,3-Dichloropropene	0.480	0.510	-6.3	81	0.00
54 T	cis-1,3-Dichloropropene	0.531	0.557	-4.9	82	0.00
55 T	1,1,2-Trichloroethane	0.323	0.342	-5.9	83	0.00
56 T	Ethyl methacrylate	0.471	0.488	-3.6	79	0.00
57 T	1,3-Dichloropropane	0.546	0.573	-4.9	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.190	0.190	0.0	77	0.00
59 T	2-Hexanone	0.248	0.253	-2.0	80	0.00
60 T	Dibromochloromethane	0.354	0.365	-3.1	80	0.00
61 T	1,2-Dibromoethane	0.339	0.340	-0.3	79	0.00
62 S	4-Bromofluorobenzene	0.379	0.389	-2.6	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.334	0.321	3.9	85	0.00
65 PM	Chlorobenzene	1.071	1.011	5.6	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.403	-2.3	83	0.00
67 C	Ethyl Benzene	1.848	1.786	3.4#	79	0.00
68 T	m/p-Xylenes	0.704	0.686	2.6	79	0.00
69 T	o-Xylene	0.691	0.668	3.3	78	0.00
70 T	Styrene	1.089	1.103	-1.3	79	0.00
71 P	Bromoform	0.254	0.261	-2.8	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	4.728	4.246	10.2	79	0.00
74 T	N-amyl acetate	1.557	1.422	8.7	78	0.00
75 P	1,1,2,2-Tetrachloroethane	1.500	1.366	8.9	82	0.00
76 T	1,2,3-Trichloropropane	1.182	1.223	-3.5	92	0.00
77 T	Bromobenzene	1.098	0.990	9.8	77	0.00
78 T	n-propylbenzene	5.137	4.793	6.7	81	0.00
79 T	2-Chlorotoluene	3.316	2.961	10.7	81	0.00
80 T	1,3,5-Trimethylbenzene	3.828	3.509	8.3	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.385	0.394	-2.3	85	0.00
82 T	4-Chlorotoluene	3.095	2.904	6.2	80	0.00
83 T	tert-Butylbenzene	3.328	2.849	14.4	76	0.00
84 T	1,2,4-Trimethylbenzene	3.711	3.473	6.4	81	0.00
85 T	sec-Butylbenzene	4.355	3.697	15.1	74	0.00
86 T	p-Isopropyltoluene	3.534	3.042	13.9	74	0.00
87 T	1,3-Dichlorobenzene	1.857	1.671	10.0	77	0.00
88 T	1,4-Dichlorobenzene	1.791	1.655	7.6	79	0.00
89 T	n-Butylbenzene	2.788	2.363	15.2	73	0.00
90 T	Hexachloroethane	0.627	0.554	11.6	75	0.00
91 T	1,2-Dichlorobenzene	1.814	1.697	6.4	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.234	0.225	3.8	81	0.00
93 T	1,2,4-Trichlorobenzene	0.646	0.558	13.6	65	0.00
94 T	Hexachlorobutadiene	0.376	0.263	30.1#	67	0.00
95 T	Naphthalene	2.175	1.754	19.4	61	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061523\
Data File : VN078220.D
Acq On : 15 Jun 2023 11:33
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 16 06:02:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.723	0.608	15.9	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6