


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061620\
Data File : VN061830.D
Acq On : 16 Jun 2020 16:32
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Title : SW846 8260
Last Update : Wed Jun 17 01:19:44 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3285

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	4512	PASS
75	95	30	60	49.5	12387	PASS
95	95	100	100	100.0	25040	PASS
96	95	5	9	6.9	1725	PASS
173	174	0.00	2	1.7	325	PASS
174	95	50	100	78.1	19565	PASS
175	174	5	9	7.3	1422	PASS
176	174	95	101	97.3	19040	PASS
177	176	5	9	7.1	1360	PASS