

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061621\
 Data File : VN067364.D
 Acq On : 16 Jun 2021 15:29
 Operator : JC/MD
 Sample : VN0616WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0616WBL01

Quant Time: Jun 17 04:19:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N061121W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 12 09:13:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.659	128	63316	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	390571	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.749	117	349420	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	202190	30.39	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.30%
60) 4-Bromofluorobenzene	12.736	95	182027	26.98	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.93%
63) Toluene-d8	10.446	98	509983	30.60	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061621\
Data File : VN067364.D
Acq On : 16 Jun 2021 15:29
Operator : JC/MD
Sample : VN0616WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0616WBL01

Quant Time: Jun 17 04:19:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N061121W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 12 09:13:15 2021
Response via : Initial Calibration

