

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061622\
 Data File : VN072902.D
 Acq On : 17 Jun 2022 00:54
 Operator : JC\MD
 Sample : N3202-06 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-28DL

Quant Time: Jun 17 06:08:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061622W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 17 06:04:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	499626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	844940	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	736440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	290062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	448300	48.566	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.140%
35) Dibromofluoromethane	7.951	113	366306	47.726	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.460%
50) Toluene-d8	10.375	98	990122	32.957	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	65.920%
62) 4-Bromofluorobenzene	12.669	95	386357	37.073	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	74.140%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.293	59	99662	41.467	ug/l	99
16) Acetone	4.240	43	18818	4.039	ug/l #	80
19) Methyl tert-butyl Ether	5.540	73	142592	5.594	ug/l	98
22) Diisopropyl ether	6.422	45	7688	0.294	ug/l #	78
25) 2-Butanone	7.275	43	13211	1.780	ug/l	92
40) Benzene	8.386	78	98091	2.937	ug/l	99
52) Toluene	10.439	92	180945	8.372	ug/l	99
67) Ethyl Benzene	11.780	91	41204	1.202	ug/l	94
68) m/p-Xylenes	11.886	106	241904	18.371	ug/l	97
69) o-Xylene	12.216	106	259820	19.232	ug/l	100
80) 1,3,5-Trimethylbenzene	12.992	105	166756	7.522	ug/l	98
84) 1,2,4-Trimethylbenzene	13.304	105	218052	10.231	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061622\
Data File : VN072902.D
Acq On : 17 Jun 2022 00:54
Operator : JC\MD
Sample : N3202-06 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-28DL

Quant Time: Jun 17 06:08:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061622W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 17 06:04:26 2022
Response via : Initial Calibration

