

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061622\
 Data File : VN072906.D
 Acq On : 17 Jun 2022 02:31
 Operator : JC\MD
 Sample : N3339-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Jun 17 06:09:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061622W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 17 06:04:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	407787	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	633193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	558011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	237853	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	345343	45.838	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.680%
35) Dibromofluoromethane	7.951	113	287435	49.974	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.940%
50) Toluene-d8	10.375	98	984006	43.707	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.420%
62) 4-Bromofluorobenzene	12.669	95	359089	45.979	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061622\
Data File : VN072906.D
Acq On : 17 Jun 2022 02:31
Operator : JC\MD
Sample : N3339-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Jun 17 06:09:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061622W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 17 06:04:26 2022
Response via : Initial Calibration

