

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061622\
 Data File : VN072895.D
 Acq On : 16 Jun 2022 22:03
 Operator : JC\MD
 Sample : N3202-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-31

Quant Time: Jun 17 06:07:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061622W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 17 06:04:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	660241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	1100286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	969613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	392643	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	494465	40.536	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	81.080%
35) Dibromofluoromethane	7.951	113	429829	43.006	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	86.020%
50) Toluene-d8	10.375	98	1401559	35.826	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	71.660%
62) 4-Bromofluorobenzene	12.669	95	537300	39.592	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	79.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061622\
Data File : VN072895.D
Acq On : 16 Jun 2022 22:03
Operator : JC\MD
Sample : N3202-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-31

Quant Time: Jun 17 06:07:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061622W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 17 06:04:26 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.016 min Scan# 1042
Delta R.T. -0.000 min
Lab File: VN072895.D
Acq: 16 Jun 2022 22:03




Instrument :
MSVOA_N
ClientSampleId :
MW-31


Tgt Ion:168 Resp: 660241
Ion Ratio Lower Upper
168 100
99 56.7 42.3 63.5



#33
1,2-Dichloroethane-d4
Concen: 40.536 ug/l
RT: 8.369 min Scan# 1102
Delta R.T. 0.000 min
Lab File: VN072895.D
Acq: 16 Jun 2022 22:03

Tgt Ion: 65 Resp: 494465
Ion Ratio Lower Upper
65 100
67 51.9 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.898 min Scan# 1192

Delta R.T. 0.000 min

Lab File: VN072895.D

Acq: 16 Jun 2022 22:03

Instrument :

MSVOA_N

ClientSampleId :

MW-31





Tgt Ion:114 Resp: 1100286

Ion Ratio Lower Upper

114 100

63 19.9 0.0 40.0

88 15.6 0.0 32.6



#35

Dibromofluoromethane

Concen: 43.006 ug/l

RT: 7.951 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VN072895.D

Acq: 16 Jun 2022 22:03




Tgt Ion:113 Resp: 429829


Ion Ratio Lower Upper

113 100

111 103.5 81.4 122.0

192 19.8 17.0 25.6









#50
Toluene-d8
Concen: 35.826 ug/l
RT: 10.375 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN072895.D
Acq: 16 Jun 2022 22:03




Instrument :
MSVOA_N
ClientSampleId :
MW-31


Tgt Ion: 98 Resp: 1401559
Ion Ratio Lower Upper
98 100
100 65.8 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 39.592 ug/l
RT: 12.669 min Scan# 1833
Delta R.T. 0.006 min
Lab File: VN072895.D
Acq: 16 Jun 2022 22:03

Tgt Ion: 95 Resp: 537300
Ion Ratio Lower Upper
95 100
174 76.9 0.0 167.6
176 74.4 0.0 163.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.680 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN072895.D

Acq: 16 Jun 2022 22:03

Instrument :

MSVOA_N

ClientSampleId :

MW-31





Tgt Ion:117 Resp: 969613

Ion Ratio Lower Upper

117 100

82 54.2 44.2 66.2

119 33.5 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.610 min Scan# 1993

Delta R.T. 0.000 min

Lab File: VN072895.D

Acq: 16 Jun 2022 22:03

Tgt Ion:152 Resp: 392643

Ion Ratio Lower Upper

152 100

115 59.1 28.9 86.7

150 156.7 0.0 356.2

