



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061622\
 Data File : VN072886.D
 Acq On : 16 Jun 2022 17:38
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061622W.M
 Title : SW846 8260
 Last Update : Fri Jun 17 06:04:26 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	13308	PASS
75	95	30	60	49.6	35437	PASS
95	95	100	100	100.0	71509	PASS
96	95	5	9	7.0	5020	PASS
173	174	0.00	2	1.1	601	PASS
174	95	50	100	77.5	55427	PASS
175	174	5	9	7.2	4016	PASS
176	174	95	101	96.9	53709	PASS
177	176	5	9	6.7	3619	PASS