

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
 Data File : VN078253.D
 Acq On : 16 Jun 2023 12:26
 Operator : JC\MD
 Sample : 03299-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-061523

Quant Time: Jun 17 02:21:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	456383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	821582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	737249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	219545	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	320266	52.632	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.260%
35) Dibromofluoromethane	8.171	113	287021	55.252	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.500%
50) Toluene-d8	10.571	98	971870	48.530	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.060%
62) 4-Bromofluorobenzene	12.853	95	300431	48.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.520%
Target Compounds						
60) Dibromochloromethane	11.365	129	62756	10.794	ug/l	Qvalue 98
71) Bromoform	12.582	173	48831	13.030	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
Data File : VN078253.D
Acq On : 16 Jun 2023 12:26
Operator : JC\MD
Sample : 03299-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-061523

Quant Time: Jun 17 02:21:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

