

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
 Data File : VN078255.D
 Acq On : 16 Jun 2023 13:14
 Operator : JC\MD
 Sample : 03231-11DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-3D-06122023DL

Quant Time: Jun 17 02:22:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	457041	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	838351	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	734421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	228796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	324731	53.289	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.580%
35) Dibromofluoromethane	8.177	113	290015	54.712	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.420%
50) Toluene-d8	10.571	98	984846	48.194	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.380%
62) 4-Bromofluorobenzene	12.853	95	301537	47.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.940%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.359	130	158411	26.426	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
Data File : VN078255.D
Acq On : 16 Jun 2023 13:14
Operator : JC\MD
Sample : 03231-11DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-3D-06122023DL

Quant Time: Jun 17 02:22:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

