

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
 Data File : VN078256.D
 Acq On : 16 Jun 2023 13:41
 Operator : JC\MD
 Sample : PB153477TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153477TB

Quant Time: Jun 17 02:22:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	451816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	844073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	799997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	249396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	322470	53.530	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.060%	
35) Dibromofluoromethane	8.171	113	289632	54.269	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.540%	
50) Toluene-d8	10.571	98	1023863	49.764	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.520%	
62) 4-Bromofluorobenzene	12.853	95	331573	51.840	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.680%	
Target Compounds						
						Qvalue
16) Acetone	4.442	43	27198	15.686	ug/l	88
37) Ethyl Acetate	7.571	43	61394	10.266	ug/l	96
43) Isopropyl Acetate	8.700	43	284172	24.847	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
Data File : VN078256.D
Acq On : 16 Jun 2023 13:41
Operator : JC\MD
Sample : PB153477TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153477TB

Quant Time: Jun 17 02:22:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

