

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
 Data File : VN078257.D
 Acq On : 16 Jun 2023 14:05
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 17 02:23:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	322334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	589890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	534171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	145740	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	232650	54.134	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.260%
35) Dibromofluoromethane	8.177	113	200499	53.756	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.520%
50) Toluene-d8	10.571	98	741267	51.553	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.100%
62) 4-Bromofluorobenzene	12.853	95	202050	45.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
Data File : VN078257.D
Acq On : 16 Jun 2023 14:05
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 17 02:23:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

