

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
 Data File : VN078264.D
 Acq On : 16 Jun 2023 16:53
 Operator : JC\MD
 Sample : 03262-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WS-SOIL2-0615

Quant Time: Jun 17 02:26:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	440345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	827883	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	785089	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	256969	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	319478	54.415	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.820%			
35) Dibromofluoromethane	8.177	113	283120	54.086	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.180%			
50) Toluene-d8	10.576	98	1022019	50.645	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.300%			
62) 4-Bromofluorobenzene	12.853	95	330559	52.692	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 105.380%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	181445	107.368	ug/l	93
20) Methylene Chloride	5.289	84	22496	4.028	ug/l #	91
25) 2-Butanone	7.494	43	61566	23.691	ug/l #	77
37) Ethyl Acetate	7.577	43	68880	11.743	ug/l #	94
43) Isopropyl Acetate	8.694	43	325900	29.052	ug/l #	80
51) 4-Methyl-2-Pentanone	10.453	43	426472	72.197	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
Data File : VN078264.D
Acq On : 16 Jun 2023 16:53
Operator : JC\MD
Sample : 03262-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WS-SOIL2-0615

Quant Time: Jun 17 02:26:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

