

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
 Data File : VN078266.D
 Acq On : 16 Jun 2023 17:41
 Operator : JC\MD
 Sample : 03282-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-NEX-001-G

Quant Time: Jun 17 02:26:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	422161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	812097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	732214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	228153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	313608	55.716	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.440%
35) Dibromofluoromethane	8.177	113	275685	53.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.380%
50) Toluene-d8	10.571	98	953149	48.151	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.300%
62) 4-Bromofluorobenzene	12.853	95	293896	47.758	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.520%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	52046	32.124	ug/l	95
20) Methylene Chloride	5.283	84	67510	12.610	ug/l	93
37) Ethyl Acetate	7.583	43	31607	5.493	ug/l	97
43) Isopropyl Acetate	8.694	43	281587	25.590	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061623\
Data File : VN078266.D
Acq On : 16 Jun 2023 17:41
Operator : JC\MD
Sample : 03282-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-NEX-001-G

Quant Time: Jun 17 02:26:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

