

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087023.D
 Acq On : 16 Jun 2025 12:22
 Operator : JC\MD
 Sample : Q2299-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE117D1-20250609

Quant Time: Jun 17 07:21:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	248893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	491697	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	447651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	215929	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	173132	51.954	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.900%		
35) Dibromofluoromethane	8.177	113	147882	50.750	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.500%		
50) Toluene-d8	10.570	98	601581	52.151	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.300%		
62) 4-Bromofluorobenzene	12.847	95	219116	51.127	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.260%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.394	101	3512	1.294	ug/l #	79
12) 1,1-Dichloroethene	4.371	96	4876	1.760	ug/l	91
27) cis-1,2-Dichloroethene	7.494	96	3491	0.947	ug/l	83
38) Carbon Tetrachloride	8.371	117	4417	1.036	ug/l	94
44) Trichloroethene	9.359	130	317312	94.249	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	1574	0.471	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087023.D
Acq On : 16 Jun 2025 12:22
Operator : JC\MD
Sample : Q2299-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE117D1-20250609

Quant Time: Jun 17 07:21:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

