

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087024.D
 Acq On : 16 Jun 2025 12:43
 Operator : JC\MD
 Sample : Q2299-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT191D2-20250609

Quant Time: Jun 17 07:21:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	210130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	419506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	373250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	177105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	149231	53.043	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.080%	
35) Dibromofluoromethane	8.177	113	127825	51.416	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.840%	
50) Toluene-d8	10.565	98	513664	52.192	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.380%	
62) 4-Bromofluorobenzene	12.847	95	184236	50.386	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.780%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.412	101	7016	3.063	ug/l	96
12) 1,1-Dichloroethene	4.371	96	1707	0.730	ug/l #	94
27) cis-1,2-Dichloroethene	7.506	96	1276	0.410	ug/l	80
38) Carbon Tetrachloride	8.377	117	1227	0.337	ug/l	91
44) Trichloroethene	9.359	130	183196	63.777	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087024.D
Acq On : 16 Jun 2025 12:43
Operator : JC\MD
Sample : Q2299-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT191D2-20250609

Quant Time: Jun 17 07:21:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

