

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087027.D
 Acq On : 16 Jun 2025 13:47
 Operator : JC\MD
 Sample : Q2299-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW10A-MW01S-20250609

Quant Time: Jun 17 07:22:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	204665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	406666	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	176246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	146180	53.346	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.700%	
35) Dibromofluoromethane	8.177	113	124238	51.551	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.100%	
50) Toluene-d8	10.570	98	500154	52.424	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.840%	
62) 4-Bromofluorobenzene	12.847	95	179098	50.527	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.060%	
Target Compounds						Qvalue
16) Acetone	4.453	43	1511	1.040	ug/l	95
22) Diisopropyl ether	6.688	45	14449	1.814	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087027.D
Acq On : 16 Jun 2025 13:47
Operator : JC\MD
Sample : Q2299-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW10A-MW01S-20250609

Quant Time: Jun 17 07:22:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

