

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087028.D
 Acq On : 16 Jun 2025 14:08
 Operator : JC\MD
 Sample : Q2299-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW10A-MW01I-20250609

Quant Time: Jun 17 07:23:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	211245	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	419153	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	374756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	175252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	148717	52.581	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.160%	
35) Dibromofluoromethane	8.177	113	127300	51.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.500%	
50) Toluene-d8	10.565	98	510077	51.871	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.740%	
62) 4-Bromofluorobenzene	12.847	95	181483	49.675	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.340%	
Target Compounds						
16) Acetone	4.447	43	1846	1.231	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087028.D
Acq On : 16 Jun 2025 14:08
Operator : JC\MD
Sample : Q2299-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW10A-MW01I-20250609

Quant Time: Jun 17 07:23:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

