

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087029.D
 Acq On : 16 Jun 2025 14:29
 Operator : JC\MD
 Sample : Q2299-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT15811-20250610

Quant Time: Jun 17 07:23:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	260360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	512497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	459773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	216825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	182557	52.370	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.740%		
35) Dibromofluoromethane	8.177	113	156413	51.499	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.000%		
50) Toluene-d8	10.565	98	627361	52.178	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 104.360%		
62) 4-Bromofluorobenzene	12.847	95	224511	50.259	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.520%		
Target Compounds						
16) Acetone	4.459	43	2105	1.139	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087029.D
Acq On : 16 Jun 2025 14:29
Operator : JC\MD
Sample : Q2299-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT15811-20250610

Quant Time: Jun 17 07:23:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

