

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087033.D
 Acq On : 16 Jun 2025 15:54
 Operator : JC\MD
 Sample : Q2299-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT17411-20250610

Quant Time: Jun 17 07:24:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	203501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	412992	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	175583	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	149110	54.726	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.460%	
35) Dibromofluoromethane	8.177	113	126731	51.780	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.560%	
50) Toluene-d8	10.565	98	506443	52.270	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.540%	
62) 4-Bromofluorobenzene	12.847	95	176642	49.071	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.140%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.371	96	2483	1.096	ug/l	88
24) 1,1-Dichloroethane	6.594	63	14741	3.235	ug/l	97
44) Trichloroethene	9.353	130	5588	1.976	ug/l	93
64) Tetrachloroethene	11.106	164	1386	0.590	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087033.D
Acq On : 16 Jun 2025 15:54
Operator : JC\MD
Sample : Q2299-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT174I1-20250610

Quant Time: Jun 17 07:24:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

