

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087034.D
 Acq On : 16 Jun 2025 16:15
 Operator : JC\MD
 Sample : Q2299-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D4-20250610

Quant Time: Jun 17 07:25:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	201975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	409116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	370070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	169597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	149258	55.194	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.380%	
35) Dibromofluoromethane	8.177	113	126336	52.108	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.220%	
50) Toluene-d8	10.571	98	503264	52.434	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	104.860%	
62) 4-Bromofluorobenzene	12.847	95	175542	49.227	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	98.460%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.401	101	957277	434.736	ug/l	Qvalue 99
27) cis-1,2-Dichloroethene	7.500	96	3374	1.128	ug/l	84
44) Trichloroethene	9.359	130	100648	35.929	ug/l	99
64) Tetrachloroethene	11.106	164	73927	31.565	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087034.D
Acq On : 16 Jun 2025 16:15
Operator : JC\MD
Sample : Q2299-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D4-20250610

Quant Time: Jun 17 07:25:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

