

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087035.D
 Acq On : 16 Jun 2025 16:37
 Operator : JC\MD
 Sample : Q2299-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D3-20250610

Quant Time: Jun 17 07:25:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	194213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	382353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	344228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	162344	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	139461	53.633	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.260%	
35) Dibromofluoromethane	8.177	113	117345	51.787	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.580%	
50) Toluene-d8	10.571	98	470679	52.472	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.940%	
62) 4-Bromofluorobenzene	12.847	95	164461	49.348	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.700%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.400	101	1794131	847.348	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	12801	4.451	ug/l	89
44) Trichloroethene	9.359	130	491001	187.546	ug/l	97
45) 1,2-Dichloropropane	9.624	63	3139	1.169	ug/l	98
64) Tetrachloroethene	11.106	164	616337	282.918	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087035.D
Acq On : 16 Jun 2025 16:37
Operator : JC\MD
Sample : Q2299-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D3-20250610

Quant Time: Jun 17 07:25:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

