

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
 Data File : VN087039.D
 Acq On : 16 Jun 2025 18:02
 Operator : JC\MD
 Sample : Q2299-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D1-20250611

Quant Time: Jun 17 07:27:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	205063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	411667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	176304	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	147287	53.645	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.300%	
35) Dibromofluoromethane	8.177	113	126902	52.017	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.040%	
50) Toluene-d8	10.571	98	506312	52.425	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.840%	
62) 4-Bromofluorobenzene	12.847	95	178049	49.621	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.240%	
Target Compounds						
						Qvalue
16) Acetone	4.459	43	1846	1.268	ug/l #	80
44) Trichloroethene	9.353	130	7823	2.775	ug/l	90
64) Tetrachloroethene	11.106	164	1026	0.437	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061625\
Data File : VN087039.D
Acq On : 16 Jun 2025 18:02
Operator : JC\MD
Sample : Q2299-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D1-20250611

Quant Time: Jun 17 07:27:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

