

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061719\
 Data File : VN056340.D
 Acq On : 17 Jun 2019 15:10
 Operator : JC/SP
 Sample : K3392-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OD-1

Quant Time: Jun 18 05:19:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:23:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	342134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	574031	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	576753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	180508	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	196806	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
35) Dibromofluoromethane	7.59	113	170464	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
50) Toluene-d8	10.09	98	718553	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.40	95	238538	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
Target Compounds						
					Qvalue	
16) Acetone	3.82	43	57128	26.875	ug/l	99
18) Methyl Acetate	4.33	43	11730	1.433	ug/l	91
20) Methylene Chloride	4.54	84	8488m	2.261	ug/l	
43) Isopropyl Acetate	8.17	43	12479	1.594	ug/l #	91
52) Toluene	10.16	92	15571	1.628	ug/l	98
95) Naphthalene	15.13	128	3144	5.006	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061719\
Data File : VN056340.D
Acq On : 17 Jun 2019 15:10
Operator : JC/SP
Sample : K3392-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OD-1

Quant Time: Jun 18 05:19:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:23:59 2019
Response via : Initial Calibration

