

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061719\
 Data File : VN056341.D
 Acq On : 17 Jun 2019 15:34
 Operator : JC/SP
 Sample : K3407-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OILY-DEBRIS

Quant Time: Jun 18 05:22:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:23:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	330340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	559330	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	578600	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	237189	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	191690	51.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.22%	
35) Dibromofluoromethane	7.59	113	166106	46.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.62%	
50) Toluene-d8	10.09	98	708966	51.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.40	95	259990	57.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.52%	
Target Compounds						
					Qvalue	
16) Acetone	3.82	43	19785	9.640	ug/l	99
17) Carbon Disulfide	4.04	76	324469	39.556	ug/l	100
20) Methylene Chloride	4.55	84	4265	1.177	ug/l	94
43) Isopropyl Acetate	8.17	43	7895	1.035	ug/l	98
51) 4-Methyl-2-Pentanone	9.99	43	61057	12.555	ug/l	98
52) Toluene	10.16	92	398066	42.714	ug/l	99
95) Naphthalene	15.13	128	3998	4.995	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061719\
Data File : VN056341.D
Acq On : 17 Jun 2019 15:34
Operator : JC/SP
Sample : K3407-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OILY-DEBRIS

Quant Time: Jun 18 05:22:36 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:23:59 2019
Response via : Initial Calibration

