

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061719\
 Data File : VN056346.D
 Acq On : 17 Jun 2019 17:33
 Operator : JC/SP
 Sample : K3154-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-061419-A

Quant Time: Jun 18 05:40:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:23:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	341671	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	578295	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	584074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	184922	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	198584	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
35) Dibromofluoromethane	7.59	113	171526	46.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.50%	
50) Toluene-d8	10.09	98	733327	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
62) 4-Bromofluorobenzene	12.40	95	241638	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
Target Compounds						
16) Acetone	3.82	43	25148	11.846	ug/l	97
20) Methylene Chloride	4.55	84	187519	50.025	ug/l	99
43) Isopropyl Acetate	8.17	43	13682	1.734	ug/l	97
95) Naphthalene	15.13	128	9699	5.700	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061719\
Data File : VN056346.D
Acq On : 17 Jun 2019 17:33
Operator : JC/SP
Sample : K3154-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-061419-A

Quant Time: Jun 18 05:40:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:23:59 2019
Response via : Initial Calibration

