

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061840.D
 Acq On : 17 Jun 2020 11:41
 Operator : JC/MD
 Sample : VN0617WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0617WBL01

Quant Time: Jun 18 04:24:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	152192	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	271336	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	251201	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97507	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	95088	46.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.18%	
35) Dibromofluoromethane	7.54	113	87103	50.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.36%	
50) Toluene-d8	10.06	98	331917	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
62) 4-Bromofluorobenzene	12.37	95	106098	45.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061840.D
Acq On : 17 Jun 2020 11:41
Operator : JC/MD
Sample : VN0617WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0617WBL01

Quant Time: Jun 18 04:24:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

