

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061843.D
 Acq On : 17 Jun 2020 13:04
 Operator : JC/MD
 Sample : L2989-03DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D1-20200609DL

Quant Time: Jun 18 04:36:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	139419	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	246121	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	240229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	106737	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	87913	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
35) Dibromofluoromethane	7.54	113	76456	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
50) Toluene-d8	10.06	98	307315	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
62) 4-Bromofluorobenzene	12.38	95	109937	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	2747	1.943	ug/l	91
27) cis-1,2-Dichloroethene	6.77	96	1517	0.808	ug/l	82
44) Trichloroethene	8.80	130	80038	44.722	ug/l	97
64) Tetrachloroethene	10.60	164	2252	1.471	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061843.D
Acq On : 17 Jun 2020 13:04
Operator : JC/MD
Sample : L2989-03DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D1-20200609DL

Quant Time: Jun 18 04:36:27 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

