

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061851.D
 Acq On : 17 Jun 2020 16:15
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 18 05:01:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	139355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	253051	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	244958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	93607	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	91229	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
35) Dibromofluoromethane	7.55	113	80351	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
50) Toluene-d8	10.06	98	317311	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
62) 4-Bromofluorobenzene	12.38	95	105007	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
Target Compounds						
44) Trichloroethene	8.80	130	1965	1.068	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061851.D
Acq On : 17 Jun 2020 16:15
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 18 05:01:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

