

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061859.D
 Acq On : 17 Jun 2020 19:27
 Operator : JC/MD
 Sample : L3003-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-0568

Quant Time: Jun 18 06:31:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	134414	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	246388	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	229506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	85349	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	89334	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	7.55	113	79169	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
50) Toluene-d8	10.06	98	301482	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	12.38	95	96071	45.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.40%	
Target Compounds						
4) Vinyl Chloride	2.16	62	10068	4.527	ug/l	95
21) trans-1,2-Dichloroethene	5.00	96	14376	9.164	ug/l	96
27) cis-1,2-Dichloroethene	6.78	96	23514	12.991	ug/l	89
44) Trichloroethene	8.80	130	15326	8.554	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061859.D
Acq On : 17 Jun 2020 19:27
Operator : JC/MD
Sample : L3003-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-0568

Quant Time: Jun 18 06:31:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

