

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061860.D
 Acq On : 17 Jun 2020 19:52
 Operator : JC/MD
 Sample : L3003-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-TB-2020-2

Quant Time: Jun 18 06:33:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	132201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	240200	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	232851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	91332	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	87039	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
35) Dibromofluoromethane	7.55	113	77894	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	10.06	98	298830	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	12.38	95	99861	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061860.D
Acq On : 17 Jun 2020 19:52
Operator : JC/MD
Sample : L3003-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-TB-2020-2

Quant Time: Jun 18 06:33:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

