

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061866.D
 Acq On : 17 Jun 2020 23:04
 Operator : JC/MD
 Sample : VN0617WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0617WBL02

Quant Time: Jun 18 06:45:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	135579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	248023	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	235636	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	90980	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	89361	49.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.28%	
35) Dibromofluoromethane	7.55	113	78289	49.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.68%	
50) Toluene-d8	10.06	98	305624	48.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.98%	
62) 4-Bromofluorobenzene	12.38	95	99880	46.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.36%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061866.D
Acq On : 17 Jun 2020 23:04
Operator : JC/MD
Sample : VN0617WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0617WBL02

Quant Time: Jun 18 06:45:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

