

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061882.D
 Acq On : 18 Jun 2020 5:29
 Operator : JC/MD
 Sample : PB129582ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129582ZHE#01

Quant Time: Jun 18 07:26:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	133815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	249960	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	252650	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	94319	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	90055	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
35) Dibromofluoromethane	7.55	113	79235	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
50) Toluene-d8	10.06	98	318925	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
62) 4-Bromofluorobenzene	12.38	95	106854	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
Target Compounds						
16) Acetone	3.78	43	6412	12.096	ug/l	94
18) Methyl Acetate	4.29	43	1812	1.187	ug/l #	69
20) Methylene Chloride	4.50	84	3643	2.109	ug/l #	95
43) Isopropyl Acetate	8.13	43	31169	10.388	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061882.D
Acq On : 18 Jun 2020 5:29
Operator : JC/MD
Sample : PB129582ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129582ZHE#01

Quant Time: Jun 18 07:26:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

