

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061720\
 Data File : VN061883.D
 Acq On : 18 Jun 2020 5:53
 Operator : JC/MD
 Sample : PB129582ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129582ZHE#02

Quant Time: Jun 18 07:28:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	135380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	253674	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	254785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	91601	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
35) Dibromofluoromethane	7.55	113	80991	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
50) Toluene-d8	10.06	98	325196	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	12.38	95	108205	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
Target Compounds						
16) Acetone	3.78	43	6889	12.845	ug/l	92
18) Methyl Acetate	4.28	43	2207	1.429	ug/l #	85
20) Methylene Chloride	4.50	84	3813	2.182	ug/l	89
43) Isopropyl Acetate	8.13	43	31629	10.387	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061720\
Data File : VN061883.D
Acq On : 18 Jun 2020 5:53
Operator : JC/MD
Sample : PB129582ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129582ZHE#02

Quant Time: Jun 18 07:28:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

