

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061720\
 Data File : VN061839.D
 Acq On : 17 Jun 2020 11:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 18 04:18:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	0.582	0.732	-25.8#	122	0.00
3 P	Chloromethane	0.694	0.714	-2.9	110	0.00
4 C	Vinyl Chloride	0.827	0.891	-7.7#	113	0.00
5 T	Bromomethane	0.635	0.618	2.7	111	0.00
6 T	Chloroethane	0.560	0.586	-4.6	112	0.00
7 T	Trichlorofluoromethane	0.988	1.112	-12.6	120	0.00
8 T	Diethyl Ether	0.332	0.361	-8.7	117	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.507	0.564	-11.2	123	0.00
10 T	Methyl Iodide	0.547	0.683	-24.9	129	0.00
11 T	Tert butyl alcohol	0.093	0.082	11.8	101	0.00
12 CM	1,1-Dichloroethene	0.508	0.556	-9.4#	119	0.00
13 T	Acrolein	0.055	0.053	3.6	113	0.00
14 T	Allyl chloride	0.674	0.729	-8.2	117	0.00
15 T	Acrylonitrile	0.230	0.226	1.7	106	0.00
16 T	Acetone	0.198	0.227	-14.6	138	0.00
17 T	Carbon Disulfide	1.486	1.606	-8.1	117	0.00
18 T	Methyl Acetate	0.570	0.538	5.6	106	0.00
19 T	Methyl tert-butyl Ether	1.691	1.785	-5.6	115	0.00
20 T	Methylene Chloride	0.645	0.628	2.6	115	0.00
21 T	trans-1,2-Dichloroethene	0.584	0.623	-6.7	117	0.00
22 T	Diisopropyl ether	1.542	1.630	-5.7	118	0.00
23 T	Vinyl Acetate	1.213	1.272	-4.9	114	0.00
24 P	1,1-Dichloroethane	1.009	1.059	-5.0	118	0.00
25 T	2-Butanone	0.289	0.296	-2.4	113	0.00
26 T	2,2-Dichloropropane	0.849	0.936	-10.2	126	0.00
27 T	cis-1,2-Dichloroethene	0.673	0.714	-6.1	118	0.00
28 T	Bromochloromethane	0.449	0.491	-9.4	129	0.00
29 T	Tetrahydrofuran	0.190	0.181	4.7	105	0.00
30 C	Chloroform	1.079	1.098	-1.8#	115	0.00
31 T	Cyclohexane	0.884	0.916	-3.6	120	0.00
32 T	1,1,1-Trichloroethane	0.919	0.964	-4.9	117	0.00
33 S	1,2-Dichloroethane-d4	0.671	0.655	2.4	114	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.320	0.339	-5.9	116	0.00
36 T	1,1-Dichloropropene	0.456	0.513	-12.5	117	0.00
37 T	Ethyl Acetate	0.356	0.370	-3.9	107	0.00
38 T	Carbon Tetrachloride	0.452	0.500	-10.6	117	0.00
39 T	Methylcyclohexane	0.491	0.602	-22.6	123	0.00
40 TM	Benzene	1.381	1.518	-9.9	116	0.00
41 T	Methacrylonitrile	0.152	0.155	-2.0	115	0.00
42 TM	1,2-Dichloroethane	0.471	0.506	-7.4	113	0.00
43 T	Isopropyl Acetate	0.600	0.642	-7.0	110	0.00
44 TM	Trichloroethene	0.364	0.404	-11.0	119	0.00
45 C	1,2-Dichloropropane	0.340	0.376	-10.6#	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061720\
 Data File : VN061839.D
 Acq On : 17 Jun 2020 11:04
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 18 04:18:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.239	0.259	-8.4	112	0.00
47 T	Bromodichloromethane	0.490	0.537	-9.6	115	0.00
48 T	Methyl methacrylate	0.273	0.285	-4.4	110	0.00
49 T	1,4-Dioxane	0.006	0.007	-16.7	108	0.00
50 S	Toluene-d8	1.258	1.343	-6.8	118	0.00
51 T	4-Methyl-2-Pentanone	0.355	0.371	-4.5	108	0.00
52 CM	Toluene	0.868	0.996	-14.7#	116	0.00
53 T	t-1,3-Dichloropropene	0.520	0.600	-15.4	120	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.651	-15.4	119	0.00
55 T	1,1,2-Trichloroethane	0.341	0.378	-10.9	115	0.00
56 T	Ethyl methacrylate	0.456	0.518	-13.6	112	0.00
57 T	1,3-Dichloropropane	0.578	0.627	-8.5	113	0.00
58 T	2-Chloroethyl Vinyl ether	0.180	0.175	2.8	100	0.00
59 T	2-Hexanone	0.253	0.278	-9.9	112	0.00
60 T	Dibromochloromethane	0.381	0.419	-10.0	115	0.00
61 T	1,2-Dibromoethane	0.356	0.382	-7.3	111	0.00
62 S	4-Bromofluorobenzene	0.431	0.475	-10.2	120	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.319	0.371	-16.3	122	0.00
65 PM	Chlorobenzene	1.055	1.158	-9.8	119	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.419	-13.9	117	0.00
67 C	Ethyl Benzene	1.769	2.057	-16.3#	119	0.00
68 T	m/p-Xylenes	0.682	0.812	-19.1	120	0.00
69 T	o-Xylene	0.651	0.766	-17.7	118	0.00
70 T	Styrene	1.090	1.312	-20.4	118	0.00
71 P	Bromoform	0.268	0.302	-12.7	113	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.553	4.004	-12.7	120	0.00
74 T	N-amyl acetate	1.140	1.249	-9.6	114	0.00
75 P	1,1,2,2-Tetrachloroethane	1.155	1.121	2.9	111	0.00
76 T	1,2,3-Trichloropropane	0.982	0.897	8.7	106	0.00
77 T	Bromobenzene	0.896	0.969	-8.1	119	0.00
78 T	n-propylbenzene	4.054	4.635	-14.3	122	0.00
79 T	2-Chlorotoluene	2.435	2.665	-9.4	120	0.00
80 T	1,3,5-Trimethylbenzene	2.940	3.376	-14.8	120	0.00
81 T	trans-1,4-Dichloro-2-butene	0.360	0.380	-5.6	116	0.00
82 T	4-Chlorotoluene	2.516	2.823	-12.2	121	0.00
83 T	tert-Butylbenzene	2.492	2.785	-11.8	118	0.00
84 T	1,2,4-Trimethylbenzene	2.902	3.374	-16.3	119	0.00
85 T	sec-Butylbenzene	3.262	3.787	-16.1	123	0.00
86 T	p-Isopropyltoluene	2.907	3.490	-20.1	123	0.00
87 T	1,3-Dichlorobenzene	1.635	1.800	-10.1	119	0.00
88 T	1,4-Dichlorobenzene	1.673	1.811	-8.2	120	0.00
89 T	n-Butylbenzene	2.427	2.901	-19.5	125	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061720\
Data File : VN061839.D
Acq On : 17 Jun 2020 11:04
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 18 04:18:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.532	0.587	-10.3	118	0.00
91 T	1,2-Dichlorobenzene	1.574	1.711	-8.7	116	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.188	0.191	-1.6	108	0.00
93 T	1,2,4-Trichlorobenzene	0.743	0.887	-19.4	124	0.00
94 T	Hexachlorobutadiene	0.293	0.327	-11.6	119	0.00
95 T	Naphthalene	2.140	2.437	-13.9	116	0.00
96 T	1,2,3-Trichlorobenzene	0.736	0.850	-15.5	120	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6