

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061721\
 Data File : VN067383.D
 Acq On : 17 Jun 2021 17:50
 Operator : JC/MD
 Sample : PB137151ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137151ZHE#01

Manual Integrations
 APPROVED

MMDadoda
 6/18/2021 9:13:15 AM

Quant Time: Jun 18 05:11:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	325421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	645226	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	585626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	210884	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	319550	49.777	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.560%		
35) Dibromofluoromethane	8.024	113	203762	47.483	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.960%		
50) Toluene-d8	10.443	98	855142	51.992	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.980%		
62) 4-Bromofluorobenzene	12.733	95	311513	45.893	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.780%		
Target Compounds					Qvalue	
16) Acetone	4.304	43	28443m	11.546	ug/l	
20) Methylene Chloride	5.101	84	23809	4.645	ug/l	87
43) Isopropyl Acetate	8.566	43	60620	4.407	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061721\
Data File : VN067383.D
Acq On : 17 Jun 2021 17:50
Operator : JC/MD
Sample : PB137151ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137151ZHE#01

Manual Integrations
APPROVED

MMDadoda
6/18/2021 9:13:15 AM

Quant Time: Jun 18 05:11:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

