

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061721\
 Data File : VN067386.D
 Acq On : 17 Jun 2021 19:03
 Operator : JC/MD
 Sample : M2746-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NEW-DORP-2

Quant Time: Jun 18 05:12:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	312781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	648656	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	562194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	199546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	306714	49.708	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.420%
35) Dibromofluoromethane	8.024	113	196989	45.662	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.320%
50) Toluene-d8	10.443	98	837866	50.672	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.340%
62) 4-Bromofluorobenzene	12.733	95	296909	43.511	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.020%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	121368	51.256	ug/l	92
18) Methyl Acetate	4.867	43	13729	2.391	ug/l #	77
20) Methylene Chloride	5.101	84	26007	5.279	ug/l	95
43) Isopropyl Acetate	8.563	43	56056	4.054	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061721\
Data File : VN067386.D
Acq On : 17 Jun 2021 19:03
Operator : JC/MD
Sample : M2746-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NEW-DORP-2

Quant Time: Jun 18 05:12:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

