

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061724\
 Data File : VN082612.D
 Acq On : 17 Jun 2024 16:34
 Operator : JC\MD
 Sample : P2894-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Jun 18 04:59:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	131807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	240619	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	233842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96896	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	114547	54.193	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.380%
35) Dibromofluoromethane	8.171	113	84093	51.517	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.040%
50) Toluene-d8	10.571	98	308524	49.980	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.960%
62) 4-Bromofluorobenzene	12.853	95	100932	47.426	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.860%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061724\
Data File : VN082612.D
Acq On : 17 Jun 2024 16:34
Operator : JC\MD
Sample : P2894-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Jun 18 04:59:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. 0.001 min
Lab File: VN082612.D
Acq: 17 Jun 2024 16:34




Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4


Tgt Ion:168 Resp: 131807
Ion Ratio Lower Upper
168 100
99 60.2 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 54.193 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.000 min
Lab File: VN082612.D
Acq: 17 Jun 2024 16:34

Tgt Ion: 65 Resp: 114547
Ion Ratio Lower Upper
65 100
67 51.0 0.0 104.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.106 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082612.D
Acq: 17 Jun 2024 16:34




Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4


Tgt Ion:114 Resp: 240619
Ion Ratio Lower Upper
114 100
63 23.9 0.0 44.6
88 15.2 0.0 31.4



#35
Dibromofluoromethane
Concen: 51.517 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.006 min
Lab File: VN082612.D
Acq: 17 Jun 2024 16:34

Tgt Ion:113 Resp: 84093
Ion Ratio Lower Upper
113 100
111 104.3 82.4 123.6
192 18.5 14.9 22.3









#50
Toluene-d8
Concen: 49.980 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN082612.D
Acq: 17 Jun 2024 16:34




Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4


Tgt Ion: 98 Resp: 308524
Ion Ratio Lower Upper
98 100
100 63.8 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 47.426 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN082612.D
Acq: 17 Jun 2024 16:34

Tgt Ion: 95 Resp: 100932
Ion Ratio Lower Upper
95 100
174 79.1 0.0 159.2
176 74.7 0.0 147.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.871 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN082612.D

Acq: 17 Jun 2024 16:34

Instrument :

MSVOA_N

ClientSampleId :

Storage-Blank-WATER-REF-4





Tgt Ion:117 Resp: 233842

Ion Ratio Lower Upper

117 100

82 60.2 47.5 71.3

119 30.9 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.000 min

Lab File: VN082612.D

Acq: 17 Jun 2024 16:34

Tgt Ion:152 Resp: 96896

Ion Ratio Lower Upper

152 100

115 59.5 30.6 91.6

150 156.6 0.0 348.6

